

CERTIFIKAT O KALIBRACIJI / CALIBRATION CERTIFICATE

stran / page 1 od 2

Št. certifikata / Certificate no. 12861-LTM-21

Naročnik
Applicant 5Labs d.o.o.
Legen 112
2380 SLOVENJ GRADEC

Instrument
Instrument Črtno merilo

Proizvajalec MITUTOYO

Manufacturer

Tip (0-50/0,1) mm

Type

Serijska številka 25921

Serial no.

Kalibracijska metoda Primerjava izmerjene vrednosti z dolžinskim etalonom - SOP 21
Calibration method

Pogoji okolja Temperatura ($20 \pm 0,1$) °C
Environmental conditions

Čas in mesto kalibracije Od 22. 10. 2021 do 22. 10. 2021 LTM Maribor
Calibration period

Rezultati Rezultati kalibracije so podani na strani 2
Results

Negotovost Razširjena negotovost meritve je podana kot standardna negotovost pomnožena s faktorjem $k = 2$, ki ustreza intervalu zaupanja 95 %.
Uncertainty
The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which corresponds to coverage probability approximately 95 %.

Pri meritvah so bili uporabljeni etaloni, ki so sledljivi na primarni etalon za dolžino.

The measurements have been executed using standards, which are traceable to primary standard of length.

Izvedel / Performed by

Digitaly
signed by dr.
Rok Klobučar
dr. Rok Klobučar

Odobril / Approved by

Digitaly
signed by
Bojan Ačko
prof. dr. Bojan Ačko

Datum / Date

2. 11. 2021

CERTIFIKAT O KALIBRACIJI / CALIBRATION CERTIFICATE

stran / page 2 od 2

Št. certifikata / Certificate no. 12861-LTM-21

Merilni rezultati Measurement results

Zap. št.	Nazivna dolžina / Nominal length L_n v mm	Razbirek na etalonu / Reference value L_e v mm	Razbirek na merilu / Instrument indication R v mm	Odstopanje / Deviation σ v μm
1	0	0,00000	0	0,00
2	5	4,99951	5	0,49
3	10	9,99936	10	0,64
4	15	14,99948	15	0,52
5	20	19,99886	20	1,14
6	25	24,99846	25	1,54
7	30	29,99862	30	1,38
8	35	34,99920	35	0,80
9	40	39,99891	40	1,09
10	45	44,99884	45	1,16
11	50	49,99968	50	0,32

Izmerjene dolžine veljajo za temperaturo 20 °C in za stanje merila med kalibracijo. Za korekcijo merilnih rezultatov na temperaturo 20 °C je bila uporabljena temperaturna razteznost $\alpha = 6 \cdot 10^{-6} \text{ K}^{-1}$.

Measured lengths are valid for temperature 20 °C and for condition of gauge as it was during calibration. For the correction of the results to temp. 20 °C a temperature expansion coefficient $\alpha = 6 \cdot 10^{-6} \text{ K}^{-1}$ was used.

Razlaga rezultata

Results explanation

$$\sigma = R - L_e$$

σ – odstopanje / deviation

R – razbirek na merilu / instrument indication

L_e – razbirek na etalonu / reference value

Merilna negotovost

Uncertainty of measurement

$$U = 0,15 \mu\text{m} + 1,1 \cdot 10^{-6} \cdot L_n$$

Uporabljeni etaloni in merilna oprema

Applied standards and measurement equipment

Etalon Standard	Proizvajalec Manufacturer	Tip Type	Oznaka Identification	Št. kalibracije Calibration No.
Laserski interferometer	HP	5528A	2532A02451	12423-LTM-20

Opomba

Remark