

KALIBRACIJSKI CERTIFIKAT

Calibration certificate

NAROČNIK
Customer 5Labs d.o.o.
Legen 112, 2380 Slovenj Gradec

LASTNIK
Owner 5Labs d.o.o.
Legen 112, 2380 Slovenj Gradec

MERILO
Object Etaloni debeline • Thickness standards

IDENTIFIKACIJA
Identification 01806075 PROIZVAJALEC
Manufacturer /

TIP
Type / MERILNI OBSEG
Meas. range (0,025 - 2) mm



**SLOVENSKA
AKREDITACIJA**
SIST EN ISO/IEC 17025
LK-034

Slovenska akreditacija je podpisnica večstrankarskih sporazumov o priznavanju akreditacijskih organov z Evropsko akreditacijo (EA - MLA) in Mednarodnim združenjem za akreditacijo laboratorijev (ILAC - MRA). Slovenian Accreditation is signatory to the multilateral agreements on recognition of accreditation bodies with the European Accreditation (EA - MLA) and International Laboratory Accreditation Cooperation (ILAC - MRA).

KRAJ KALIBRACIJE
Place of calibration Slovenj Gradec

DATUM PREJEMA
Date of receipt 18. 11. 2023

DATUM KALIBRACIJE
Date of calibration 18. 11. 2023

OPOMBE
Notes

ODOBRIL
Approved by Petra Oprešnik

Dovoljeno je razmnoževanje le celotnega certifikata. Verodostojnost podpisa je mogoče preveriti v elektronski obliki certifikata. Only the reproduction of the complete certificate is allowed. Signature validity can be verified in electronic version of certificate

1. KALIBRACIJSKI POSTOPEK • Calibration procedure

Kalibracija je narejena v skladu z internim kalibracijskim postopkom KL-5Labs-19 in obsega določitev pogreškov debelinskega merila.

Calibration is made in accordance with the internal calibration procedure KL-5Labs-19 and comprises the determination of thickness gauge errors.

2. KALIBRACIJSKI POGOJI • Calibration conditions

Temperatura • Temperature $(20 \pm 1) ^\circ\text{C}$

3. SLEDLJIVOST • Traceability

Identifikacije pri kalibraciji uporabljenih etalonov ali referenčnih merilnikov
Standards or reference measuring device identification used at calibration

1812			
------	--	--	--

Ta kalibracijski certifikat dokumentira sledljivost do (inter) nacionalnih etalonov v skladu z mednarodnim sistemom merskih enot (SI).

This calibration certificate documents the traceability to (inter) national standards, which realize the units of measurement according to the International System of Units (SI).

Kalibracijski certifikati zgoraj navedene opreme uporabljene pri kalibraciji so javno objavljeni na naši spletni strani pod https://www.5labs.si/index.php/open_documents

The calibration certificates of the above mentioned equipment used at calibration are publicly available on our website at https://www.5labs.si/index.php/open_documents

4. MERILNA NEGOTOVOST • Measuring uncertainty

$$U = 0,3 \mu\text{m} + 0,0000071 \cdot L$$

Merilna negotovost je podana kot standardna negotovost meritve pomnožena s faktorjem $k = 2$, ki pri normalni porazdelitvi ustrza verjetnosti 95%. Standardna merilna negotovost je določena v skladu s publikacijo EA-4/02.

The specification indicates the expanded measuring uncertainty resulting from multiplication of standard measuring uncertainty by the factor $k = 2$. It was determined in conformity with EA-4/02. The values of the measurement parameter lie within the specified range with a probability of 95%.

5. RAZLAGA REZULTATOV • Result explanation

Pogrešek = Izmerjena vrednost - Referenčna vrednost

Error = Measured value - Reference value

Povprečen pogrešek = povprečje treh pogreškov

Average error = the average of the tree errors

Podani merilni rezultati in pripadajoča merilna negotovost se nanašajo samo na to kalibrano merilo. Izmerjene vrednosti veljajo v času meritev ne zagotavljajo dolgotrajne stabilnosti.

The measurement results and uncertainties quoted refer only to this calibrated gauge. The measurement results are valid at the time of measurement do not guarantee long-term stability.

6. MERILNI REZULTATI • Measurement results

Nominalna vrednost L • Nominal value L [mm]	Pogrešek 1 • Error 1 [mm]	Pogrešek 2 • Error 2 [mm]	Pogrešek 3 • Error 3 [mm]	Povp. pogrešek • Avr. error [mm]
0,0220	0,0002	0,0007	0,0007	0,0005
0,0490	-0,0002	-0,0004	-0,0005	-0,0004
0,0990	0,0008	0,0006	0,0004	0,0006
0,2500	-0,0007	-0,0005	-0,0001	-0,0004
0,4540	0,0008	0,0010	0,0006	0,0008
0,4720	-0,0011	0,0009	0,0017	0,0005
1,9870	-0,0007	-0,0004	-0,0002	-0,0004