

# KALIBRACIJSKI CERTIFIKAT

## Calibration certificate

NAROČNIK  
Costumer            5Labs d.o.o.  
                          Legen 112, 2380 Slovenj Gradec

LASTNIK  
Owner                5Labs d.o.o.  
                          Legen 112, 2380 Slovenj Gradec

MERILO  
Object              Dvotočkovni mikrometer za zunanje merjenje • A two-point micrometer  
                          for external measurement

IDENTIFIKACIJA  
Identification        67395718                    PROIZVAJALEC  
    Manufacturer                Mitutoyo

TIP  
Type                 293-240-30                    MERILNI OBSEG  
    Meas. range                (0 - 25) mm

LOČLJIVOST  
Resolution          0,001 mm                    VRSTA MIKROMETRA  
    Type of micrometer        Digitalno • Digital



**SLOVENSKA  
AKREDITACIJA**  
SIST EN ISO/IEC 17025  
**LK-034**

Slovenska akreditacija je podpisnica večstrankarskih sporazumov o priznavanju akreditacijskih organov z Evropsko akreditacijo (EA - MLA) in Mednarodnim združenjem za akreditacijo laboratoriјev (ILAC - MRA). Slovenian Accreditation is signatory to the multilateral agreements on recognition of accreditation bodies with the European Accreditation (EA - MLA) and International Laboratory Accreditation Cooperation (ILAC - MRA).

KRAJ KALIBRACIJE        Slovenj Gradec  
Place of calibration

DATUM KALIBRACIJE      18. 11. 2023  
Date of calibration

DATUM PREJEMA          18. 11. 2023  
Date of receipt

OPOMBE  
Notes

ODOBRIL                 Petra Oprešnik  
Approved by

Dovoljeno je razmnoževanje le celotnega certifikata. Verodostojnost podpisa je mogoče preveriti v elektronski obliki certifikata. Only the reproduction of the complete certificate is allowed. Signature validity can be verified in electronic version of certificate.

## 1. KALIBRACIJSKI POSTOPEK • Calibration procedure

Kalibracija mikrometra je narejena v skladu z interno proceduro KL-5Labs-3 in standardom SIST EN ISO 3611:2011. Obsega pregled funkcionalnosti, določitev pogreškov kazanja mikrometra in morebitnih merilnih vstavkov.

Calibration of the micrometer is made in accordance with internal procedure KL-5Labs-3 and standard SIST EN ISO 3611:2011. Includes review of functionality, determination of indication errors and micrometer measuring inserts, if they exist.

## 2. KALIBRACIJSKI POGOJI • Calibration conditions

Temperatura • Temperature  $(20 \pm 1)^\circ\text{C}$

## 3. SLEDLJIVOST • Traceability

Identifikacije pri kalibraciji uporabljenih etalonov

Standards identification used at calibration

1200907    

Ta kalibracijski certifikat dokumentira sledljivost do (inter) nacionalnih etalonov v skladu z mednarodnim sistemom merskih enot (SI).

This calibration certificate documents the traceability to (inter) national standards, which realize the units of measurement according to the International System of Units (SI).

Kalibracijski certifikati zgoraj navedene opreme uporabljene pri kalibraciji so javno objavljeni na naši spletni strani pod [https://www.5labs.si/index.php/open\\_documents](https://www.5labs.si/index.php/open_documents)

The calibration certificates of the above mentioned equipment used at calibration are publicly available on our website at [https://www.5labs.si/index.php/open\\_documents](https://www.5labs.si/index.php/open_documents)

## 4. MERILNA NEGOTOVOST • Measuring uncertainty

$$U = 1,6 \mu\text{m} + 0,00001 * L$$

Merilna negotovost je podana kot standardna negotovost meritve pomnožena s faktorkem  $k = 2$ , ki pri normalni porazdelitvi ustreza verjetnosti 95%. Standardna merilna negotovost je določena v skladu s publikacijo EA-4/02.

The specification indicates the expanded measuring uncertainty resulting from multiplication of standard measuring uncertainty by the factor  $k = 2$ . It was determined in conformity with EA-4/02. The values of the measurement parameter lie within the specified range with a probability of 95%.

## 5. RAZLAGA REZULTATOV • Result explanation

Pogrešek = Izmerjena vrednost - Referenčna vrednost

Error = Measured value - Reference value

Podani merilni rezultati in pripadajoča merilna negotovost se nanašajo samo na to kalibrirano merilo. Izmerjene vrednosti veljajo v času meritev in ne zagotavljajo dolgotrajne stabilnosti.


The measurement results and uncertainties quoted refer only to this calibrated gauge. The measurement results are valid at the time of measurement and do not guarantee long-term stability.

## 6. MERILNI REZULTATI • Measurement results

Odstopanje ničelne lege pred nastavitevijo • Deviation before the zero position adjustment [mm]

0 mm

Ref. dolžina L • Ref. length L [mm]	Izmerjena vrednost • Measured value [mm]	Pogrešek • Error [µm]	Toleranca • Tolerance (DIN 863-1:2017) [µm]
0,000	0,000	0	4
2,500	2,500	0	4
5,100	5,100	0	4
7,700	7,700	0	4
10,300	10,300	0	4
12,900	12,900	0	4
15,000	15,000	0	4
17,600	17,600	0	4
20,200	20,200	0	4
22,800	22,800	0	4
25,000	25,000	0	4



UGOTOVITEV  
Findings

Pogreški merila ustrezajo zahtevam podanim v DIN 863-1:2017, brez upoštevanja meritve negotovosti. • The measurement errors meet the requirements given in DIN 863-1: 2017, without taking measurement uncertainty into account.